

Boule Con-Diff

CONTROL

IVD



Control	Lot	
Tri-Level	22401-1K	2024-05-30
Low	22401-11	2024-05-30
Normal	22401-12	2024-05-30
High	22401-13	2024-06-04

Open vial stability 14 days ¹

SWELAB

CONTROL		L	
LOT		22401-11	
Swelab Alfa/Alfa Plus		Mean ²	± Range ³
RBC	10 ¹² /L	2.23	0.12
MCV	fL	68.4	5.0
RDW	%	14.0	4.0
Hct	%	15.3	2.5
Plt	10 ⁹ /L	81	15
MPV	fL	12.3	1.5
WBC	10 ⁹ /L	3.4	0.3
Hgb	g/dL	5.6	0.3
MCH	pg	25.1	1.5
MCHC	g/dL	36.6	3.0
Lymph#	10 ⁹ /L	1.6	0.4
Mid#	10 ⁹ /L	0.1	0.3
Gran#	10 ⁹ /L	1.7	0.8
Lymph	%	48.0	7.0
Mid	%	3.1	5.0
Gran	%	48.9	7.0

CONTROL		N	
LOT		22401-12	
		Mean ²	± Range ³
		4.22	0.18
		80.3	5.0
		13.2	4.0
		33.9	3.0
		233	30
		12.5	1.5
		8.4	0.6
		11.9	0.4
		28.2	1.5
		35.1	3.0
		4.2	0.6
		0.3	0.5
		3.9	1.0
		50.1	5.0
		3.3	5.0
		46.6	5.0

CONTROL		H	
LOT		22401-13	
		Mean ²	± Range ³
		5.10	0.22
		93.5	5.0
		12.5	4.0
		47.7	3.5
		534	60
		12.9	1.5
		20.6	1.8
		16.4	0.5
		32.2	1.5
		34.4	3.0
		13.3	1.5
		0.6	1.0
		6.7	2.5
		64.5	5.0
		3.0	5.0
		32.5	5.0

1

bg	Трайност на отворен флакон 14 дни (bg)
hr	Stabilnost otvorene bočice 14 dana (hr)
cs	Stabilita otevřené ampulky 14 dní (cs)
da	Åben rør stabilitet 14 dage (da)
et	Stabiilsus avatud viaalis 14 päevade (et)
fr	Stabilité en flacon ouvert 14 jour (fr)
de	Stabilität geöffneter Flaschen 14 tage (de)
el	Διατήρηση δείγματος μετά την αποσφράγιση 14 ημέρα (el)
it	Stabilità della fiala aperta 14 giorni (it)
lv	Atvērt flakonu stabilitāti 14 dienas (lv)
lt	Stabilumas atidarius buteliuką 14 dienas (lt)
no	Åpen rør/glass stabilitet 14 dager (no)
pl	Trwałość po otwarciu fiołki 14 Liczba dni (pl)
pt	Estabilidade após abertura do frasco 14 dias (pt)
ro	Termenul de valabilitate al fiolei desfăcute 14 zile (ro)
ru	Стабильность вскрытого флакона 14 дней (ru)
sr	Proizvod stabilan nakon otvaranja pakovanja 14 dana (sr)
sk	Stabilita otvorenej injekčnej liekovky 14 dní (sk)
sl	Stabilita otvorenej injekčnej liekovky 14 dni (sl)
es	Estabilidad de la cápsula abierta 14 días (es)
sv	Hållbarhet för öppnad flaska 14 dagar (sv)
tr	Açık şişe stabilitesi 14 gün (tr)

2

Средно
Sredstvo
Průměr
Middelverdi
Keskmine
Moyenne
Mittelwert
Μέση
Media
Nozīmē
Vidurkis
Middelverdi
Średnia
Média
Medie
Среднее
Srednja vrednost
Priemer
Povprečno
Media
Medelvärde
Ortalama

3

Обхват
Raspon
Rozsah
Område
Vahemik
Intervalle
Bereich
Αναμενόμενο εύρος α
Intervallo
Diapazons
Intervalas
Område
Zakres
Intervalo
Interval
диапазон
Orseg
Rozsah
Območje
Intervalo
Intervall
Aralığı

For Swelab Alfa / Alfa Plus with barcode reader

Alfa:

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Alfa Plus:

1. In the Main Menu go to "Quality Control" "Input Assays"
2. Scan in each blood calibrator / control column by column, sequence 1 to 9

Instr.: Swelab Alfa
Lot: 22401-11
Prod.: Con-Diff Low
ID: 2240111+
Exp.: 30-May-2024

Instr.: Swelab Alfa
Lot: 22401-12
Prod.: Con-Diff Norm
ID: 2240112+
Exp.: 30-May-2024

Instr.: Swelab Alfa
Lot: 22401-13
Prod.: Con-Diff High
ID: 2240113+
Exp.: 4-Jun-2024

