

LOT: 12305-52  
EXP: 2023-10-22  
Open Vial Stability: 16 days

EXIGO		Mean	± Range	Expected Range	
WBC	10 <sup>9</sup> /L	8.3	0.6	7.7	8.9
Hct	%	29.4	2.5	26.9	31.9
MCV	fL	70.3	5.0	65.3	75.3
RBC	10 <sup>12</sup> /L	4.18	0.18	4.00	4.36
Hgb	g/dL	10.0	0.4	9.6	10.4
MCH	pg	23.9	1.5	22.4	25.4
MCHC	g/dL	34.0	2.2	31.8	36.2
PLT	10 <sup>9</sup> /L	218	30	188	248

This assay sheet is also available on :

[www.boule.com](http://www.boule.com)

## For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo  
Lot: 12305-52  
Prod.: Vet-Con Norm  
ID: 1230552+  
Exp.: 22-Oct-2023

