

LOT: 12311-52
EXP: 2024-04-21
Open Vial Stability: 16 days

EXIGO		Mean	± Range	Expected Range	
WBC	$10^9/L$	8.6	0.6	8.0	9.2
Hct	%	28.3	2.5	25.8	30.8
MCV	fL	67.5	5.0	62.5	72.5
RBC	$10^{12}/L$	4.19	0.18	4.01	4.37
Hgb	g/dL	10.1	0.4	9.7	10.5
MCH	pg	24.1	1.5	22.6	25.6
MCHC	g/dL	35.7	2.2	33.5	37.9
PLT	$10^9/L$	219	30	189	249

This assay sheet is also available on :
www.boule.com

For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo
Lot: 12311-52
Prod.: Vet-Con Norm
ID: 1231152+
Exp.: 21-Apr-2024

