

LOT: 12402-52
EXP: 2024-07-21
Open Vial Stability: 16 days

EXIGO		Mean	± Range	Expected Range	
WBC	$10^9/L$	8.5	0.6	7.9	9.1
Hct	%	29.6	2.5	27.1	32.1
MCV	fL	71.0	5.0	66.0	76.0
RBC	$10^{12}/L$	4.17	0.18	3.99	4.35
Hgb	g/dL	10.6	0.4	10.2	11.0
MCH	pg	25.4	1.5	23.9	26.9
MCHC	g/dL	35.8	2.2	33.6	38.0
PLT	$10^9/L$	205	30	175	235

This assay sheet is also available on :
www.boule.com

For EXIGO/EXIGO EOS with barcode reader

1. In the Main Menu go to "Q/C Menu" "Enter Con/Cal" "Control/Calibrator Assay Value Input"
2. Scan in each blood calibrator/control column by column, sequence 1 to 9

Instr.: Exigo
Lot: 12402-52
Prod.: Vet-Con Norm
ID: 1240252+
Exp.: 21-Jul-2024

